

# AUTHORIZATION TO MARK LISTING REPORT

320/VD

This authorizes the manufacturer to apply the *ETL* and *CETL* mark to classified products when made in accordance with the accompanying descriptions and drawings under the conditions set forth in the Certification Agreement herein:

**Applicant:** WR Medical Electronics  
123-T North 2<sup>nd</sup> Street  
Stillwater, MN 55082

**Contact:** Mr. Jack Blais: Phone: (651) 430-8474,  
FAX: (651) 439-9733

**Manufacturer:** Same as Applicant

**Reference Report No.:** J20053402.003

**Product Covered:** Q-Sweat, model 5188

**Description:** Quantitative Sweat Measurement System

**Standards:** Standards for Medical Electrical Equipment, Part 1:  
General Requirements for Safety, UL 2601-1, Second  
Edition: 1997, and CAN/CSA-C22.2 No. 601.1-M90.

This procedure, with all revisions, etc., is the property of Intertek Testing Services and is intended solely for the guidance of the listee and the representative of Intertek Testing Services, and is not transferable.

Issued by: Intertek Testing Services, 24 Groton Avenue, Cortland, NY 13045-2014  
USA



Authorized by:

*William T. Starr*  
William T. Starr  
Certification Manager

Date:

*5/23/02*

*(Signature)*

Control No: *3025752*