

AUTHORIZATION TO MARK LISTING REPORT

This authorizes the *application of the Listing Marks shown below to the Product Covered* (also to the multiple listee model *identified* on the correlation page of the Listing Report where applicable) when made in accordance with the *Description* and under the conditions set forth in the Certification Agreement and Listing Report:

Applicant: WR Medical Electronics
123-T North 2nd Street
Stillwater, MN 55082

Manufacturer: WR Medical Electronics
123 North 2nd Street
Stillwater, MN 55082

Contact: Mr. Tim Wall: Phone: (651) 430-1200,
FAX: (651) 430-8449

Party Authorized To Apply Mark: *(Same as Manufacturer)*

Report No.: 3058136.001

Report Issuing Office: *Oakdale, Minnesota*

Product Covered: Facial Nerve Stimulator Silverstein, Model S8

Description: Facial Nerve Stimulator

Standards: Standards for Medical Electrical Equipment, Part 1: General Requirements for Safety, UL 60601-1, First Edition: 2003, and CAN/CSA-C22.2 No. 601.1-M90.

This document is the property of Intertek Testing Services and is not transferable. Only the Applicant may reproduce this document. The certification mark(s) may be applied only at the above noted location of the Party Authorized to Apply Mark.



Authorized by: William T. Starr Date: May 5, 2004
William T. Starr, Certification Manager

Control Number: 3058136.001

This document supercedes all previous Authorizations to Mark for the noted Report Number.

Intertek Testing Services NA Inc.
165 Main Street, Corland, NY 13045
Telephone 800-345-3851 or 607-753-6711 Fax 607-756-6699